



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), गुवाहाटी
All India Institute of Medical Sciences, Guwahati
स्वास्थ्य और परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक वैधानिक निकाय
(A statutory body under the aegis of Ministry of Health and Family Welfare, GoI)
Changsari, Assam PIN- 781101

4-177/2023-24/AIIMS/GHY/PROC-Pedia/2647

Date:22.02.2024

Corrigendum-01

GEM Bid No.:GEM/2023/B/4276081 Dated 15-12-2023

Name of Equipment: - EMG NCV EP Machine (01 No.)

Amendment in Specifications

| SI No | Specification mentioned in the GeM bid document | Amended as |
|-------|--|--|
| 1 | Standards: Performance parameters: Number of Channels: 4 | The Amplifier should have at least 4 optically isolated channels with 24-bit ADC, CMRR should be >110 dB, Input impedance should be >100 Mega Ohms & Noise should be <0.6 μ V RMS |
| 2 | Additional Points to be incorporated | <ol style="list-style-type: none">1. System should have Dual Electrical Stimulator with option for Constant Current Electrical stimulator with both current (0 to 100mA) stimulation facility.2. Facilities to perform the multi-MUP analysis, Spike Triggered EMG. Single motor unit potential analysis using trigger/delayline/averager, Mutichannel EMG recordings, Mutichannel recordings with amplitude trigger/delayline/Averager, EMG recordings using single or multiple channels, Motor unit number estimation, Heart rate variability, Quantitative EMG test like Multiple Motor unit analysis, turns and amplitude analysis. |

The bid submission date is hereby extended for another seven (07) days on GeM Portal.

Extended date may please be referred on GeM Portal against the tender no.

SD/-

Administrative Officer
AIIMS, Guwahati